Keystone

FOR LABORATORY RESEARCH USE ONLY.

keystonecompounds.com

CERTIFICATE NO.
PL–26–00223

VERIFICATION KEY
LHD769QR

Scan or enter both at prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE · ISSUED FOR KEYSTONE

Kisspeptin-10 10 mg

Lot **PL-2610-109** · Lyophilized powder · Label claim **10 mg per vial** · Distributed as **Kisspeptin-10 10 mg**

TESTED IN-HOUSE BY PRYMALAB

In-house analytical result, not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL–2610–109	2026-10-05	2026-10-06	2026-10-21

RETEST / EXPIRY	UNITS IN LOT / SAMPLED	STORAGE	CERTIFICATE TYPE
2028-10-05	368 / 3	−20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND
Kisspeptin-10
KP-10; human metastin 45–54

CAS NUMBER
374675-21-5

MOLECULAR FORMULA
$C_{63}H_{83}N_{17}O_{14}$

THEORETICAL MASS (AVG.)
1,302.44 Da

SEQUENCE
YNWNSFGLRF–NH2

PURITY · RP-HPLC, AREA %
99.61%
Specification ≥ 99.0% • main peak at 12.85 min

IDENTITY · OBSERVED MASS
1,302.49 Da
Theoretical 1,302.44 Da • +0.05 Da (38 ppm)

OVERALL CONCLUSION
PASS
Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da	1,302.49 Da	**PASS**
Purity	RP-HPLC, UV 214 nm, area%	≥ 99.0%	99.61%	**PASS**
Largest impurity	RP-HPLC, area%	≤ 1.0%	0.17% (RT 11.60)	**PASS**
Water content	Karl Fischer	≤ 5.0%	1.44%	**PASS**
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.177 EU/mg	**PASS**
Sterility	USP <71>, 14-day	No growth	No growth	**PASS**

1 HPLC CHROMATOGRAM

Data file: PL-2610-109-HPLC.raw · Column C18, 150 × 4.6 mm · UV 214 nm



Peak plot drawn from the integrated peak table (area %, log scale); it is not the instrument trace, which is retained in PL-2610-109-HPLC.raw. † Nominal reference retention time for the assigned impurity; the main-peak RT is observed in this lot.

For laboratory research use only. Not for human or veterinary use.

2 HPLC METHOD · PL-2610-109-HPLC.RAW

INSTRUMENT
Waters Alliance e2695

DATA SYSTEM
Empower 3

DETECTOR
UV absorbance

DETECTION WAVELENGTH
214 nm

COLUMN
C18, 150 × 4.6 mm, 3.5 µm

MOBILE PHASE A
0.1% TFA in water

MOBILE PHASE B
0.1% TFA in MeCN

GRADIENT
5–95% B over 25 min (linear)

FLOW RATE
1.0 mL/min

COLUMN TEMPERATURE
40 °C

INJECTION VOLUME
10 µL

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
25.0	5	95

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	11.60 †	—	0.17
2	**12.85**	**—**	**99.61**
3	13.40 †	—	0.09
4	14.10 †	—	0.13
Total		—	100.00

Peaks: 1 tryptophan oxidation product; 2 main peak (Kisspeptin-10); 3 deamidation product; 4 truncated sequence · Area % from the integrated peak table (absolute areas are not part of the lot-release export) · † nominal reference RT for the assigned impurity · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT
Waters SQD

IONIZATION / MODE
ESI, positive

REFERENCE M/Z (CHARGE)
652.23 (2+) · 435.15 (3+)

DECONVOLUTED MASS
1,302.49 Da

THEORETICAL MASS
1,302.44 Da

DIFFERENCE
+0.05 Da (38 ppm)

DATA FILE
PL-2610-109-MS.d

ACCEPTANCE
Within ± 1.0 Da

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content (Karl Fischer)	PrymaLab Analytical	2026-10-06	—
Endotoxin (LAL, kinetic)	PrymaLab Analytical	2026-10-06	—
Sterility (USP <71>, day-14 read-out)	PrymaLab Analytical	2026-10-20	—

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds
Analyst · performed the testing
Date: 2026-10-21

S. Calloway
Reviewer · checked against raw data
Date: 2026-10-21

M. Phelps
Lab Director · approved for release
Date: 2026-10-21

Raw data retained at: Empower 3, project PL-2610, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. This copy is issued for Keystone, a distributor of the lot; the brand shown does not change the laboratory, the method or the results. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00223 and key LHD769QR at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.